

# Supplement

Title: 771 I.S. Supplement Issue: 1  
Part Number: 2567301 Issue Date: 4/07  
Print Date: Sept. 2006 Page Count: 1  
Revision/Date:

---

---

This supplement contains information necessary to ensure the accuracy of the above Instruction Sheet.

## Change #1, 40733

Under ***Battery Replacement***, add the following under the **Warning**:

 **Caution**

**To avoid battery leakage, which could damage the Meter, remove the batteries from the Meter when not used for an extended period.**

Under ***Specifications***, change the following:

From: **Storage Temperature** -25 to 70 °C

To: **Storage Temperature** -25 °C to 60 °C

From: **EMI, RFI, EMC**

Meets all applicable requirements in EN 61326-1

To: **EMI, RFI, EMC**

Meets all applicable requirements in EN 61326-1  
For EMC field strengths of 1 V/m up to 3 V/m, add 1 mA to specifications.